

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 03214 Calib. Date: 05/10/2024 Calib. No.: 00050
Certification File No.: 03215 Cert. Date: 05/10/2024 Cert. No.: 00044
Linearity File No.: 03216 Lin. Date: 05/10/2024 Lin. No.: 00043
Solution File No.: 03234 Soln. Date: 08/21/2024 Soln. No.: 00394
Sequential File No.: 03234 File Date: 08/21/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 01/09/2026
Solution Control Lot: 24020 Bottle No.: 1369

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:47D		
Control 1 EC	0.099%	02:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:48D		
Control 2 EC	0.098%	02:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:50D		
Control 3 EC	0.098%	02:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PEREZ

First Name: EDDIE

MI: --

Signature: _____



Badge No.: 27237

Date: 08/21/2024