

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	03238	Calib. Date:	10/11/2024
Certification File No.:	03239	Cert. No.:	00045
Linearity File No.:	03240	Lin. Date:	10/11/2024
Solution File No.:	03248	Soln. No.:	00398
Sequential File No.:	03248	File Date:	10/31/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0340
Solution Control Lot:	24020	Expires:	01/09/2026
		Bottle No.:	0637

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:31S		
Control 1 EC	0.099%	02:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:33S		
Control 2 EC	0.097%	02:33S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:34S		
Control 3 EC	0.097%	02:34S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARTINEZ

First Name: JUAN

MI: C

Signature: _____

Badge No.: 27223

Date: 10/31/2024