

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PENNSAUKEN TWSP. P.D. Serial No.: ARUM-0051
Calibration File No.: 03238 Calib. Date: 10/11/2024 Calib. No.: 00051
Certification File No.: 03239 Cert. Date: 10/11/2024 Cert. No.: 00045
Linearity File No.: 03240 Lin. Date: 10/11/2024 Lin. No.: 00044
Solution File No.: 03243 Soln. Date: 10/17/2024 Soln. No.: 00397
Sequential File No.: 03243 File Date: 10/17/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340
Control Solution %: 0.100% Expires: 01/09/2026
Solution Control Lot: 24020 Bottle No.: 0122

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:56D		
Control 1 EC	0.098%	15:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58D		
Control 2 EC	0.096%	15:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:59D		
Control 3 EC	0.096%	16:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PEREZ

First Name: EDDIE

MI: --

Signature: 

Badge No.: 27237

Date: 10/17/2024