

# Calibrating Unit New Standard Solution Report

## Equipment

Alcotest 7110 MKIII-C  
Location: PENNSAUKEN TWSP. P.D. Serial No.: ARUM-0051  
Calibration File No.: 03238 Calib. Date: 10/11/2024 Calib. No.: 00051  
Certification File No.: 03239 Cert. Date: 10/11/2024 Cert. No.: 00045  
Linearity File No.: 03240 Lin. Date: 10/11/2024 Lin. No.: 00044  
Solution File No.: 03263 Soln. Date: 12/01/2024 Soln. No.: 00401  
Sequential File No.: 03263 File Date: 12/01/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340  
Control Solution %: 0.100% Expires: 01/09/2026  
Solution Control Lot: 24020 Bottle No.: 0394

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:15S		
Control 1 EC	0.098%	02:15S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:16S		
Control 2 EC	0.096%	02:17S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:18S		
Control 3 EC	0.097%	02:18S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: COFFEY

First Name: JOHN

MI: R

Signature: \_\_\_\_\_



Badge No.: 225

Date: 12/01/2024