

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03238	Calib. Date: 10/11/2024
Certification File No.:	03239	Calib. No.: 00051
Linearity File No.:	03240	Cert. Date: 10/11/2024
Solution File No.:	03254	Cert. No.: 00045
Sequential File No.:	03254	Lin. Date: 10/11/2024
		Lin. No.: 00044
		Soln. Date: 11/06/2024
		Soln. No.: 00400
		File Date: 11/06/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	24020	Expires: 01/09/2026
		Bottle No.: 0649

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:19S		
Control 1 EC	0.099%	04:19S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:20S		
Control 2 EC	0.097%	04:21S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:22S		
Control 3 EC	0.097%	04:22S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARTINEZ

First Name: JUAN

MI: C

Signature: _____

Badge No.: 27223

Date: 11/06/2024