

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 03270 Calib. Date: 02/12/2025 Calib. No.: 00052  
Certification File No.: 03271 Cert. Date: 02/12/2025 Cert. No.: 00046  
Linearity File No.: 03272 Lin. Date: 02/12/2025 Lin. No.: 00045  
Solution File No.: 03276 Soln. Date: 02/28/2025 Soln. No.: 00404  
Sequential File No.: 03276 File Date: 02/28/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 01/09/2026  
Solution Control Lot: 24020 Bottle No.: 0227

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:59S		
Control 1 EC	0.098%	22:00S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:01S		
Control 2 EC	0.096%	22:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:02S		
Control 3 EC	0.096%	22:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: MARTINEZ

First Name: JUAN

MI: C

Signature: \_\_\_\_\_

Badge No.: 27223

Date: 02/28/2025