

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 03270 Calib. Date: 02/12/2025 Calib. No.: 00052  
Certification File No.: 03271 Cert. Date: 02/12/2025 Cert. No.: 00046  
Linearity File No.: 03272 Lin. Date: 02/12/2025 Lin. No.: 00045  
Solution File No.: 03280 Soln. Date: 03/27/2025 Soln. No.: 00405  
Sequential File No.: 03280 File Date: 03/27/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 07/23/2026  
Solution Control Lot: 24260 Bottle No.: 1114

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:59S		
Control 1 EC	0.099%	23:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:00S		
Control 2 EC	0.097%	00:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:02S		
Control 3 EC	0.098%	00:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: COFFEY

First Name: JOHN

MI: R

Signature: \_\_\_\_\_



Badge No.: 225

Date: 03/28/2025