

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP, P.D.	
Calibration File No.:	03270	Calib. Date: 02/12/2025
Certification File No.:	03271	Calib. No.: 00052
Linearity File No.:	03272	Cert. Date: 02/12/2025
Solution File No.:	03290	Cert. No.: 00046
Sequential File No.:	03290	Lin. Date: 02/12/2025
		Lin. No.: 00045
		Soln. Date: 05/27/2025
		Soln. No.: 00407
		File Date: 05/27/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	24260	Expires: 07/23/2026
		Bottle No.: 1278

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:53D		
Control 1 EC	0.098%	20:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	20:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55D		
Control 2 EC	0.097%	20:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:56D		
Control 3 EC	0.097%	20:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: EVANS

First Name: BRIAN

MI: W

Signature:  #193

Badge No.: 193

Date: 05/27/2025