

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 03270 Calib. Date: 02/12/2025 Calib. No.: 00052
Certification File No.: 03271 Cert. Date: 02/12/2025 Cert. No.: 00046
Linearity File No.: 03272 Lin. Date: 02/12/2025 Lin. No.: 00045
Solution File No.: 03285 Soln. Date: 05/11/2025 Soln. No.: 00406
Sequential File No.: 03285 File Date: 05/11/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 07/23/2026
Solution Control Lot: 24260 Bottle No.: 1277

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:42D		
Control 1 EC	0.100%	20:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:44D		
Control 2 EC	0.098%	20:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:45D		
Control 3 EC	0.098%	20:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARTINEZ

First Name: JUAN

MI: C

Signature: 

Badge No.: 27223

Date: 05/11/2025