

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03296	Calib. Date: 07/09/2025
Certification File No.:	03297	Calib. No.: 00053
Linearity File No.:	03298	Cert. Date: 07/09/2025
Solution File No.:	03302	Cert. No.: 00047
Sequential File No.:	03302	Lin. Date: 07/09/2025
		Lin. No.: 00046
		Soln. Date: 08/09/2025
		Soln. No.: 00410
		File Date: 08/09/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUÑ S3-0340
Solution Control Lot:	24260	Expires: 07/23/2026
		Bottle No.: 1267

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:30D		
Control 1 EC	0.100%	01:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:31D		
Control 2 EC	0.100%	01:31D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:32D		
Control 3 EC	0.100%	01:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COFFEY

First Name: JOHN

MI: R

Signature: _____

Badge No.: 225

Date: 08/09/2025