

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	Calib. Date:	07/09/2025
Calibration File No.:	03296	Calib. No.:	00053
Certification File No.:	03297	Cert. Date:	07/09/2025
Linearity File No.:	03298	Lin. Date:	07/09/2025
Solution File No.:	03309	Soln. Date:	08/24/2025
Sequential File No.:	03309	File Date:	08/24/2025
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0340
Solution Control Lot:	24260	Expires:	07/23/2026
		Bottle No.:	1080

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:01D		
Control 1 EC	0.098%	04:01D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:02D		
Control 2 EC	0.098%	04:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:03D		
Control 3 EC	0.098%	04:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	04:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: COFFEY

First Name: JOHN

MI: R

Signature: \_\_\_\_\_

Badge No.: 225

Date: 08/24/2025