

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 03296 Calib. Date: 07/09/2025 Calib. No.: 00053
Certification File No.: 03297 Cert. Date: 07/09/2025 Cert. No.: 00047
Linearity File No.: 03298 Lin. Date: 07/09/2025 Lin. No.: 00046
Solution File No.: 03314 Soln. Date: 09/25/2025 Soln. No.: 00412
Sequential File No.: 03314 File Date: 09/25/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340
Control Solution %: 0.100% Expires: 07/23/2026
Solution Control Lot: 24260 Bottle No.: 0813

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:28D		
Control 1 EC	0.101%	21:29D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29D		
Control 2 EC	0.099%	21:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:30D		
Control 3 EC	0.100%	21:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COFFEY

First Name: JOHN

MI: R

Signature: _____



Badge No.: 225

Date: 09/25/2025