

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 03329 Calib. Date: 11/04/2025 Calib. No.: 00054
Certification File No.: 03330 Cert. Date: 11/04/2025 Cert. No.: 00048
Linearity File No.: 03331 Lin. Date: 11/04/2025 Lin. No.: 00047
Solution File No.: 03347 Soln. Date: 12/24/2025 Soln. No.: 00416
Sequential File No.: 03347 File Date: 12/24/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340
Control Solution %: 0.100% Expires: 07/23/2026
Solution Control Lot: 24260 Bottle No.: 1273

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:06S		
Control 1 EC	0.100%	02:07S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:08S		
Control 2 EC	0.098%	02:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:09S		
Control 3 EC	0.098%	02:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: EVANS

First Name: BRIAN

MI: W

Signature:  #193

Badge No.: 193

Date: 12/24/2025