

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03329	Calib. Date: 11/04/2025
Certification File No.:	03330	Calib. No.: 00054
Linearity File No.:	03331	Cert. Date: 11/04/2025
Solution File No.:	03337	Cert. No.: 00048
Sequential File No.:	03337	Lin. Date: 11/04/2025
		Lin. No.: 00047
		Soln. Date: 11/27/2025
		Soln. No.: 00415
		File Date: 11/27/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	24260	Expires: 07/23/2026
		Bottle No.: 0933

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:06S		
Control 1 EC	0.101%	23:06S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07S		
Control 2 EC	0.099%	23:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:09S		
Control 3 EC	0.099%	23:09S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CARDUCCI

First Name: TYLER

MI: J

Signature: _____



Badge No.: 239

Date: 11/27/2025