

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 03329 Calib. Date: 11/04/2025 Calib. No.: 00054
Certification File No.: 03330 Cert. Date: 11/04/2025 Cert. No.: 00048
Linearity File No.: 03331 Lin. Date: 11/04/2025 Lin. No.: 00047
Solution File No.: 03351 Soln. Date: 02/08/2026 Soln. No.: 00417
Sequential File No.: 03351 File Date: 02/08/2026

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340
Control Solution %: 0.100% Expires: 07/23/2026
Solution Control Lot: 24260 Bottle No.: 1268

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:37S		
Control 1 EC	0.102%	07:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		
Control 2 EC	0.099%	07:39S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		
Control 3 EC	0.100%	07:41S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SURGNER

First Name: ASHLEY

MI: M

Signature: _____

Ashley Surgner #256

Badge No.: 27256

Date: 02/08/2026