

# Calibrating Unit

## New Standard Solution Report

### Equipment

Alcotest 7110 MKIII-C  
Location: PENNSAUKEN TWSP. P.D.  
Serial No.: ARUM-0051  
Calibration File No.: 03214 Calib. Date: 05/10/2024 Calib. No.: 00050  
Certification File No.: 03215 Cert. Date: 05/10/2024 Cert. No.: 00044  
Linearity File No.: 03216 Lin. Date: 05/10/2024 Lin. No.: 00043  
Solution File No.: 03224 Soln. Date: 06/01/2024 Soln. No.: 00391  
Sequential File No.: 03224 File Date: 06/01/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 01/09/2026  
Solution Control Lot: 24020 Bottle No.: 1102

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:34D		
Control 1 EC	0.099%	04:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:36D		
Control 2 EC	0.097%	04:36D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:37D		
Control 3 EC	0.097%	04:38D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: COFFEY

First Name: JOHN

MI: R

Signature: \_\_\_\_\_



Badge No.: 225

Date: 06/01/2024