

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PENNSAUKEN TWSP. P.D.
Serial No.: ARUM-0051
Calibration File No.: 03214 Calib. Date: 05/10/2024 Calib. No.: 00050
Certification File No.: 03215 Cert. Date: 05/10/2024 Cert. No.: 00044
Linearity File No.: 03216 Lin. Date: 05/10/2024 Lin. No.: 00043
Solution File No.: 03219 Soln. Date: 05/11/2024 Soln. No.: 00390
Sequential File No.: 03219 File Date: 05/11/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 01/09/2026
Solution Control Lot: 24020 Bottle No.: 1126

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:37D		
Control 1 EC	0.098%	05:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:39D		
Control 2 EC	0.097%	05:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:40D		
Control 3 EC	0.097%	05:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARTINEZ

First Name: JUAN

MI: C

Signature: _____

Badge No.: 27223

Date: 05/11/2024