

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 03155 Calib. Date: 12/04/2023 Calib. No.: 00049  
Certification File No.: 03156 Cert. Date: 12/04/2023 Cert. No.: 00043  
Linearity File No.: 03157 Lin. Date: 12/04/2023 Lin. No.: 00042  
Solution File No.: 03213 Soln. Date: 05/09/2024 Soln. No.: 00388  
Sequential File No.: 03213 File Date: 05/09/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 01/09/2026  
Solution Control Lot: 24020 Bottle No.: 1120

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:26D		
Control 1 EC	0.103%	17:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:28D		
Control 2 EC	0.101%	17:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:29D		
Control 3 EC	0.100%	17:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: SEXTON

First Name: RYAN

MI: G

Signature: \_\_\_\_\_

*RYAN* *#206*

Badge No.: 206

Date: 05/09/2024