

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 03214 Calib. Date: 05/10/2024 Calib. No.: 00050  
Certification File No.: 03215 Cert. Date: 05/10/2024 Cert. No.: 00044  
Linearity File No.: 03216 Lin. Date: 05/10/2024 Lin. No.: 00043  
Solution File No.: 03231 Soln. Date: 07/14/2024 Soln. No.: 00393  
Sequential File No.: 03231 File Date: 07/14/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 01/09/2026  
Solution Control Lot: 24020 Bottle No.: 1257

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:54D		
Control 1 EC	0.099%	09:55D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		
Control 2 EC	0.097%	09:56D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57D		
Control 3 EC	0.097%	09:58D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: SEXTON

First Name: RYAN

MI: G

Signature: \_\_\_\_\_

*PH* #206

Badge No.: 206

Date: 07/14/2024